   This is done using the Q330S. The data logger should be powered on – and have been on for a few minutes and the BALER44 status lights should be off

- open the USB compartment
- change toggle to AFMT   (lift toggle switch)
- insert a USB stick which has a blank FAT32 filesystem on it
- push ATTN
            status should show "power" and "fmt mode"
            blue "media" light will turn on, "fault" light will flash, "status" will show 2 short flashes

- WAIT for "media" light to go out and for "status" to show long – short - short
- push ATTN again - "shutdown" light will turn on
- WAIT for ALL BALER 44 lights to go out

There should be the following on the disk (when viewed with any computer)
   cont
   data
   format prevent
   recover
   sdata
   wfdisc